



NEON SENSOR COMMAND, CONTROL AND CONFIGURATION (C3) DOCUMENT: BIOLOGICAL TEMPERATURE

| PREPARED BY | ORGANIZATION | DATE |
|------------------|--------------|------------|
| Mike SanClements | FIU | 03/20/2012 |
| Hank Loescher | FIU | 02/23/2012 |

| APPROVALS | ORGANIZATION | APPROVAL DATE |
|---------------|--------------|---------------|
| Kate Thibault | SCI | 07/07/2022 |
| | | |

| RELEASED BY | ORGANIZATION | RELEASE DATE |
|----------------|--------------|--------------|
| Tanisha Waters | CM | 07/07/2022 |

See configuration management system for approval history.



Change Record

| REVISION | DATE | ECO # | DESCRIPTION OF CHANGE |
|----------|------------|-----------|---|
| A | 06/22/2012 | ECO-00417 | Initial Release |
| B | 07/07/2022 | ECO-06849 | <ul style="list-style-type: none">Revised logo and fine print |



TABLE OF CONTENTS

1 DESCRIPTION..... 1

1.1 Purpose.....1

1.2 Scope.....1

2 RELATED DOCUMENTS AND ACRONYMS..... 2

2.1 Applicable Documents2

2.2 Reference Documents2

2.3 Acronyms2

2.4 Verb Convention.....2

3 INTRODUCTION..... 3

4 OVERVIEW OF SENSOR CONFIGURATION 4

5 COMMAND AND CONTROL..... 5

5.1 Error Handling.....5

6 APPENDIX & BIBLIOGRAPHY 6



1 DESCRIPTION

1.1 Purpose

This document specifies the command, control, and configuration details for operating a NEON sensor used for instrumental observations. It includes a detailed discussion of all necessary requirements for operational control parameters, conditions/constraints, set points, and any necessary error handling. All Level 0 Data Products generated by the sensor should be identified. The raw data are compensated by the DAS, but received at HQ for further processing as LO unfiltered and uncorrected data product until its associated algorithms are applied to produce a QA/QC'd L1 data product in Standard Scientific Units.

1.2 Scope

The expectation is that the Apogee Instruments SI-111 IR Radiometer w/ Thermistor Custom Coefficient (NEON P/N: 0303220001; no firmware required) will be used to make the measurements of biological temperature (AD [01]). The reference document for the Apogee Instruments SI-111 IR Radiometer w/ Thermistor Custom Coefficient is RD [03].

This document specifies the command, control, and configuration that is needed for operating this sensor. It does not provide implementation details, except for cases where these stem directly from the sensor conditions as described here.



| | | |
|--|------------------------|------------------|
| Title: NEON Command, Control, and Configuration (C3): Biological Temperature | | Date: 07/07/2022 |
| NEON Doc. #: NEON.DOC.000417 | Author: M. SanClements | Revision: B |

2 RELATED DOCUMENTS AND ACRONYMS

2.1 Applicable Documents

| | | |
|---------|-----------------|--|
| AD [01] | NEON.DOC.000001 | NEON Observatory Design (NOD) Requirements |
| AD [02] | NEON.DOC.000291 | NEON Configured Sensor List |
| AD [03] | NEON.DOC.005003 | NEON Scientific Data Products Catalog |
| AD [04] | NEON.DOC.005005 | NEON Level 0 Data Products Catalog |
| AD [05] | NEON.DOC.XXXXXX | Biological Temperature ATBD (TBW) |

2.2 Reference Documents

| | | |
|---------|--|------------------------|
| RD [01] | NEON.DOC.000008 | NEON Acronym List |
| RD [02] | NEON.DOC.000243 | NEON Glossary of Terms |
| RD [03] | Campbell Scientific, Inc., Apogee Instruments, Inc. (2010). Instruction Manual: SI-111 Precision Infrared Radiometer. Campbell Scientific, Inc. (CSI) 815 West 1800 North Logan, Utah 84321 United States. | |

2.3 Acronyms

| Acronym | Explanation |
|----------------|--|
| ATBD | Algorithm Theoretical Basis Document |
| C ³ | Command, Control, and Configuration Document |
| SOP | Standard Operating Procedures |
| QA/QC | Quality Assurance/Quality Control |
| TIS | Terrestrial Instrument System |
| L0 | Level 0 |
| L1 | Level 1 |
| ENG | NEON Engineering group |
| CI | NEON Cyberinfrastructure group |
| DPS | NEON Data Products group |
| CVAL | NEON Calibration, Validation, and Audit Laboratory |

2.4 Verb Convention

“Shall” is used whenever a statement expresses a convention that is binding. The verbs “should” and “may” express non-mandatory provisions. “Will” is used to express a declaration of purpose on the part of the design activity.



3 INTRODUCTION

The sensor configuration and sensor command and control described here are related to the biological temperature data product (FIU.0.0005.001) and sensor body temperature data product (FIU.0.0005.002) (AD[04]). A description of how sensor readings shall be converted to biological temperature in degrees Celsius is presented in the associated ATBD (AD[04]).



4 OVERVIEW OF SENSOR CONFIGURATION

The infrared broadband radiation biological temperature data product shall be unfiltered and uncorrected V. Sensor body temperature will be unfiltered and uncorrected Ω .

Table 1. Sensor configuration settings.

| Parameter | Default Setting |
|---|--|
| Biological temperature: Acquisition rate | 1 Hz |
| Sensor body temperature: Acquisition rate | 1 Hz |
| Data acquired from sensor | Biological temperature (mV); Sensor body temperature (mV) |
| Measurement mode | Run |



| | | |
|---|-------------------------------|-------------------------|
| <i>Title:</i> NEON Command, Control, and Configuration (C3): Biological Temperature | | <i>Date:</i> 07/07/2022 |
| <i>NEON Doc. #:</i> NEON.DOC.000417 | <i>Author:</i> M. SanClements | <i>Revision:</i> B |

5 COMMAND AND CONTROL

5.1 Error Handling

This sensor provides no error notification.

5.2 Sensor <device> Controls Specification

This sensor has no associated devices.



| | | |
|---|-------------------------------|-------------------------|
| <i>Title:</i> NEON Command, Control, and Configuration (C3): Biological Temperature | | <i>Date:</i> 07/07/2022 |
| <i>NEON Doc. #:</i> NEON.DOC.000417 | <i>Author:</i> M. SanClements | <i>Revision:</i> B |

6 APPENDIX & BIBLIOGRAPHY

N/A