



neon
Operated by Battelle

<i>Title:</i> NEON Sensor Command, Control and Configuration (C3) Document: Soil Water Content and Soil Water Ion Content		<i>Date:</i> 05/26/2022
<i>NEON Doc. #:</i> NEON.DOC.000613	<i>Author:</i> E. Ayres	<i>Revision:</i> E

NEON SENSOR COMMAND, CONTROL AND CONFIGURATION (C3) DOCUMENT: SOIL WATER CONTENT AND SOIL WATER ION CONTENT

PREPARED BY	ORGANIZATION	DATE
Edward Ayres	FIU	12/10/2015

APPROVALS	ORGANIZATION	APPROVAL DATE
Kate Thibault	SCI	05/26/2022

RELEASED BY	ORGANIZATION	RELEASE DATE
Tanisha Waters	CM	05/26/2022

See configuration management system for approval history.

The National Ecological Observatory Network is a project solely funded by the National Science Foundation and managed under cooperative agreement by Battelle. Any opinions, findings, and conclusions or recommendations expressed in this material are those of the author(s) and do not necessarily reflect the views of the National Science Foundation.



neon
Operated by Battelle

Title: NEON Sensor Command, Control and Configuration (C3) Document: Soil Water Content and Soil Water Ion Content

Date: 05/26/2022

NEON Doc. #: NEON.DOC.000613

Author: E. Ayres

Revision: E

Change Record

REVISION	DATE	ECO #	DESCRIPTION OF CHANGE
A	04/30/2013	ECO-00634	Initial Release
B	2/11/2015	ECO-02519	Values for calibration coefficients have been added for circumstances where soil water content sensors may need to be deployed prior to the site- and depth-specific coefficients being known
C	1/13/2016	ECO-03567	Updated to new template, updated DP numbers to new format
D	10/22/2020	ECO-06535	Changed soil water content calibration coefficients to manufacturer's default values for all future deployments.
E	05/26/2022	ECO-06832	<ul style="list-style-type: none">• Revised NEON logo and fine print



TABLE OF CONTENTS

1	DESCRIPTION.....	1
1.1	Purpose.....	1
1.2	Scope	1
2	RELATED DOCUMENTS AND ACRONYMS.....	4
2.1	Applicable Documents	4
2.2	Reference Documents	4
2.3	Acronyms	4
3	ASSEMBLY, SOIL WATER CONTENT PROFILE, POPULATION, 2 METER INTRODUCTION (CF00810000).....	5
4	ASSEMBLY, SOIL WATER CONTENT PROFILE, POPULATION, 2 METER OVERVIEW OF SENSOR CONFIGURATION (CF00810000)	6
4.1	TriSCAN Sensors.....	6
4.2	RS-232 Interface and Other Assembly Components.....	6
5	ASSEMBLY, SOIL WATER CONTENT PROFILE, POPULATION, 2 METER COMMAND AND CONTROL (CF00810000).....	7
5.1	Error Handling	7
5.2	Sensor Controls Specification.....	7
6	ASSEMBLY, SOIL WATER CONTENT PROFILE, POPULATION, 2 METER, 1 SENSOR INTRODUCTION (CF00810010).....	8
7	ASSEMBLY, SOIL WATER CONTENT PROFILE, POPULATION, 2 METER, 1 SENSOR OVERVIEW OF SENSOR CONFIGURATION (CF00810010)	9
7.1	TriSCAN Sensors.....	9
7.2	RS-232 Interface and Other Assembly Components.....	9
8	ASSEMBLY, SOIL WATER CONTENT PROFILE, POPULATION, 2 METER, 1 SENSOR COMMAND AND CONTROL (CF00810010)	10
8.1	Error Handling	10
8.2	Sensor Controls Specification.....	10
9	ASSEMBLY, SOIL WATER CONTENT PROFILE, POPULATION, 2 METER, 2 SENSORS INTRODUCTION (CF00810020).....	11
10	ASSEMBLY, SOIL WATER CONTENT PROFILE, POPULATION, 2 METER, 2 SENSORS OVERVIEW OF SENSOR CONFIGURATION (CF00810020)	12



10.1	TriSCAN Sensors.....	12
10.2	RS-232 Interface and Other Assembly Components.....	12
11	ASSEMBLY, SOIL WATER CONTENT PROFILE, POPULATION, 2 METER, 2 SENSORS COMMAND AND CONTROL (CF00810020).....	13
11.1	Error Handling	13
11.2	Sensor Controls Specification.....	13
12	ASSEMBLY, SOIL WATER CONTENT PROFILE, POPULATION, 2 METER, 3 SENSORS INTRODUCTION (CF00810030)	14
13	ASSEMBLY, SOIL WATER CONTENT PROFILE, POPULATION, 2 METER, 3 SENSORS OVERVIEW OF SENSOR CONFIGURATION (CF00810030)	15
13.1	TriSCAN Sensors.....	15
13.2	RS-232 Interface and Other Assembly Components.....	15
14	ASSEMBLY, SOIL WATER CONTENT PROFILE, POPULATION, 2 METER, 3 SENSORS COMMAND AND CONTROL (CF00810030).....	16
14.1	Error Handling	16
14.2	Sensor Controls Specification.....	16
15	ASSEMBLY, SOIL WATER CONTENT PROFILE, POPULATION, 2 METER, 4 SENSORS INTRODUCTION (CF00810040)	17
16	ASSEMBLY, SOIL WATER CONTENT PROFILE, POPULATION, 2 METER, 4 SENSORS OVERVIEW OF SENSOR CONFIGURATION (CF00810040)	18
16.1	TriSCAN Sensors.....	18
16.2	RS-232 Interface and Other Assembly Components.....	18
17	ASSEMBLY, SOIL WATER CONTENT PROFILE, POPULATION, 2 METER, 4 SENSORS COMMAND AND CONTROL (CF00810040).....	19
17.1	Error Handling	19
17.2	Sensor Controls Specification.....	19
18	ASSEMBLY, SOIL WATER CONTENT PROFILE, POPULATION, 2 METER, 6 SENSORS INTRODUCTION (CF00810060)	20
19	ASSEMBLY, SOIL WATER CONTENT PROFILE, POPULATION, 2 METER, 6 SENSORS OVERVIEW OF SENSOR CONFIGURATION (CF00810060)	21
19.1	TriSCAN Sensors.....	21
19.2	RS-232 Interface and Other Assembly Components.....	21



20 ASSEMBLY, SOIL WATER CONTENT PROFILE, POPULATION, 2 METER, 6 SENSORS COMMAND AND CONTROL (CF00810060).....	22
20.1 Error Handling	22
20.2 Sensor Controls Specification.....	22
21 ASSEMBLY, SOIL WATER CONTENT PROFILE, POPULATION, 3 METER INTRODUCTION (CF00813000).....	23
22 ASSEMBLY, SOIL WATER CONTENT PROFILE, POPULATION, 3 METER OVERVIEW OF SENSOR CONFIGURATION (CF00813000)	24
22.1 TriSCAN Sensors.....	24
22.2 RS-232 Interface and Other Assembly Components.....	24
23 ASSEMBLY, SOIL WATER CONTENT PROFILE, POPULATION, 3 METER COMMAND AND CONTROL (CF00813000).....	25
23.1 Error Handling	25
23.2 Sensor Controls Specification.....	25
24 ASSEMBLY INTEGRATION.....	26
25 APPENDIX.....	27
25.1 List of Level 0 Data Product	27
25.2 Assembly Schematic Drawing.....	34
26 BIBLIOGRAPHY	35

LIST OF TABLES

Table 1. Sensor configuration settings.....	6
Table 2. Sensor configuration settings.....	9
Table 3. Sensor configuration settings.....	12
Table 4. Sensor configuration settings.....	15
Table 5. Sensor configuration settings.....	18
Table 6. Sensor configuration settings.....	21
Table 7. Sensor configuration settings.....	24
Table 8. List of Level 0 data product associated with DPName: Soil water content and salinity	27



Title: NEON Sensor Command, Control and Configuration (C3) Document: Soil Water Content and Soil Water Ion Content		Date: 05/26/2022
NEON Doc. #: NEON.DOC.000613	Author: E. Ayres	Revision: E

1 DESCRIPTION

1.1 Purpose

This document specifies the command, control, and configuration details for operating a NEON sensor used for instrumental observations. It includes a detailed discussion of all necessary requirements for operational control parameters, conditions/constraints, set points, and any necessary error handling. All Level 0 Data Products generated by the sensor should be identified.

1.2 Scope

The Soil Water Content Profile measurements shall be made using the Sensor TriSCAN Soil Water Content and Salinity Sensor (NEON P/N: 0318950000) and Sensor Accessory EnviroSCAN RS-232 Modbus Probe Interface Firmware Version 1.3.5 (NEON P/N: 0319390000), as well as the manufacturer specified rod, tube, top cap, and plastic expandable bung. Each DGD listed in the table below contains one RS-232 Modbus Probe Interface and 1, 2, 3, 4, 6, or 8 Sensor TriSCAN sensors.

This document specifies the command, control, and configuration that are needed for operating this sensor. It does not provide implementation details, except for cases where these stem directly from the sensor conditions as described here.

A complete set of the Level 0 data products generated in this document can be found in appendix.

The Soil Water Content Profile assembly will consist of following Data Generating Device (DGD) based on Data Generating Device DGD List and Hierarchies doc (AD [05]):

DGD Agile PN	DGD Agile Description
CF00810000	Assembly, Soil Water Content Profile, Population, 2 Meter
CF00810010	Assembly, Soil Water Content Profile, Population, 2 Meter, 1 Sensor
CF00810020	Assembly, Soil Water Content Profile, Population, 2 Meter, 2 Sensors
CF00810030	Assembly, Soil Water Content Profile, Population, 2 Meter, 3 Sensors
CF00810040	Assembly, Soil Water Content Profile, Population, 2 Meter, 4 Sensors
CF00810060	Assembly, Soil Water Content Profile, Population, 2 Meter, 6 Sensors
CF00813000	Assembly, Soil Water Content Profile, Population, 3 Meter

Further detailed sensor info under each DGD is as following:

1. Under CF00810000:
 - a. NEON P/N: 0318950000, Sensor TriSCAN Soil Water Content and Salinity Sensor, no firmware.
 - b. NEON P/N: 0319390000, Sensor Accessory EnviroSCAN RS-232 Modbus Probe Interface Firmware Version 1.3.5.



.....
2. Under CF00810010:

- a. NEON P/N: 0318950000, Sensor TriSCAN Soil Water Content and Salinity Sensor, no firmware.
 - b. NEON P/N: 0319390000, Sensor Accessory EnviroSCAN RS-232 Modbus Probe Interface Firmware Version 1.3.5.
-

3. Under CF00810020:

- a. NEON P/N: 0318950000, Sensor TriSCAN Soil Water Content and Salinity Sensor, no firmware.
 - b. NEON P/N: 0319390000, Sensor Accessory EnviroSCAN RS-232 Modbus Probe Interface Firmware Version 1.3.5.
-

4. Under CF00810030:

- a. NEON P/N: 0318950000, Sensor TriSCAN Soil Water Content and Salinity Sensor, no firmware.
 - b. NEON P/N: 0319390000, Sensor Accessory EnviroSCAN RS-232 Modbus Probe Interface Firmware Version 1.3.5.
-

5. Under CF00810040:

- a. NEON P/N: 0318950000, Sensor TriSCAN Soil Water Content and Salinity Sensor, no firmware.
 - b. NEON P/N: 0319390000, Sensor Accessory EnviroSCAN RS-232 Modbus Probe Interface Firmware Version 1.3.5.
-

6. Under CF00810060:

- a. NEON P/N: 0318950000, Sensor TriSCAN Soil Water Content and Salinity Sensor, no firmware.
 - b. NEON P/N: 0319390000, Sensor Accessory EnviroSCAN RS-232 Modbus Probe Interface Firmware Version 1.3.5.
-

7. Under CF00813000:



Title: NEON Sensor Command, Control and Configuration (C3) Document: Soil Water Content and Soil Water Ion Content		Date: 05/26/2022
NEON Doc. #: NEON.DOC.000613	Author: E. Ayres	Revision: E

- a. NEON P/N: 0318950000, Sensor TriSCAN Soil Water Content and Salinity Sensor, no firmware.
 - b. NEON P/N: 0319390000, Sensor Accessory EnviroSCAN RS-232 Modbus Probe Interface Firmware Version 1.3.5.
-

Other important parts that are not a DGD:

- c. Manufacturer specified mounting rod: NEON P/N: 0319390001, Sensor Accessory EnviroSCAN Probe 0.5m Rod
 - d. Manufacturer specified mounting rod: NEON P/N: 0319390002, Sensor Accessory EnviroSCAN Probe 1m Rod
 - e. Manufacturer specified mounting rod: NEON P/N: 0319390003, Sensor Accessory EnviroSCAN Probe 1.5m Rod
 - f. Manufacturer specified mounting rod: NEON P/N: 0319390004, Sensor Accessory EnviroSCAN Probe 2m Rod
 - g. Manufacturer specified tube: NEON P/N: 0319390005, Sensor Accessory EnviroSCAN 1m Access Tube
 - h. Manufacturer specified tube NEON P/N: 0319390006, Sensor Accessory EnviroSCAN 1.5m Access Tube
 - i. Manufacturer specified tube NEON P/N: 0319390007, Sensor Accessory EnviroSCAN 2m Access Tube
 - j. Manufacturer specified tube NEON P/N: 0319390008, Sensor Accessory EnviroSCAN 2.5m Access Tube
-



2 RELATED DOCUMENTS AND ACRONYMS

2.1 Applicable Documents

Applicable documents contain information that shall be applied in the current document. Examples are higher level requirements documents, standards, rules and regulations.

AD [01]	NEON.DOC.000001	NEON Observatory Design (NOD) Requirements
AD [02]	NEON.DOC.000291	NEON Configured Sensor List
AD [03]	NEON.DOC.005003	NEON Scientific Data Products Catalog
AD [04]	NEON.DOC.005005	NEON Level 0 Data Products Catalog
AD [05]	NEON.DOC.001104	Data Generating Device DGD List and Hierarchies
AD [06]	NEON.DOC.000007	NEON Algorithm Theoretical Basis Document (ATBD) – Soil Water Content and Soil Water Ion Content

2.2 Reference Documents

Reference documents contain information complementing, explaining, detailing, or otherwise supporting the information included in the current document.

RD [01]	NEON.DOC.000008	NEON Acronym List
RD [02]	NEON.DOC.000243	NEON Glossary of Terms
RD [03]	NEON.DOC.000007	Algorithm Theoretical Basis Document (ATBD): Soil Water Content and Soil Water Ion Content

2.3 Acronyms

Acronym	Explanation
ATBD	Algorithm Theoretical Basis Document
C ³	Command, Control, and Configuration Document
SOP	Standard Operating Procedures
QA/QC	Quality Assurance/Quality Control
TIS	Terrestrial Instrument System
L0	Level 0
L1	Level 1
ENG	NEON Engineering group
CI	NEON Cyberinfrastructure group
DPS	NEON Data Products group
CVAL	NEON Calibration, Validation, and Audit Laboratory
#CVAL	Value provided by CVAL



<i>Title:</i> NEON Sensor Command, Control and Configuration (C3) Document: Soil Water Content and Soil Water Ion Content		<i>Date:</i> 05/26/2022
<i>NEON Doc. #:</i> NEON.DOC.000613	<i>Author:</i> E. Ayres	<i>Revision:</i> E

3 ASSEMBLY, SOIL WATER CONTENT PROFILE, POPULATION, 2 METER INTRODUCTION (CF00810000)

Soil Water Content Profile subsystem data product name and number are Soil water content and salinity and NEON.DOM.SIT.DP0.00094, respectively. A description of how sensor readings shall be converted to the data product is presented in the associated ATBD (AD[06]).



4 ASSEMBLY, SOIL WATER CONTENT PROFILE, POPULATION, 2 METER OVERVIEW OF SENSOR CONFIGURATION (CF00810000)

4.1 TriSCAN Sensors

Table 1. Sensor configuration settings.

Parameter	Default Setting
Acquisition rate: Soil water content	0.1 Hz
Acquisition rate: Soil water ion content	0.1 Hz
Water content high normalization value ($SWCF_{air}$)	#CVAL
Water content low normalization value ($SWCF_{water}$)	#CVAL
Water ion content high normalization value ($SWICF_{air}$)	#CVAL
Water ion content low normalization value ($SWICF_{water}$)	#CVAL
Water content calibration coefficient A	0.1957*
Water content calibration coefficient B	0.404*
Water content calibration coefficient C	0.02852*
Ion content calibration coefficient A	1
Ion content calibration coefficient B	1
Ion content calibration coefficient C	0

* Manufacturer default values (Sentek 2003).

4.2 RS-232 Interface and Other Assembly Components

No configuration is required.



5 ASSEMBLY, SOIL WATER CONTENT PROFILE, POPULATION, 2 METER COMMAND AND CONTROL (CF00810000)

There is no command and control for this assembly.

5.1 Error Handling

N/A

5.2 Sensor Controls Specification

N/A



<i>Title:</i> NEON Sensor Command, Control and Configuration (C3) Document: Soil Water Content and Soil Water Ion Content		<i>Date:</i> 05/26/2022
<i>NEON Doc. #:</i> NEON.DOC.000613	<i>Author:</i> E. Ayres	<i>Revision:</i> E

6 ASSEMBLY, SOIL WATER CONTENT PROFILE, POPULATION, 2 METER, 1 SENSOR INTRODUCTION (CF00810010)

Soil Water Content Profile subsystem data product name and number are Soil water content and salinity and NEON.DOM.SIT.DP0.00094, respectively. A description of how sensor readings shall be converted to the data product is presented in the associated ATBD (AD[03]).



7 ASSEMBLY, SOIL WATER CONTENT PROFILE, POPULATION, 2 METER, 1 SENSOR OVERVIEW OF SENSOR CONFIGURATION (CF00810010)

7.1 TriSCAN Sensors

Table 2. Sensor configuration settings.

Parameter	Default Setting
Acquisition rate: Soil water content	0.1 Hz
Acquisition rate: Soil water ion content	0.1 Hz
Water content high normalization value ($SWCF_{air}$)	#CVAL
Water content low normalization value ($SWCF_{water}$)	#CVAL
Water ion content high normalization value ($SWICF_{air}$)	#CVAL
Water ion content low normalization value ($SWICF_{water}$)	#CVAL
Water content calibration coefficient A	0.1957*
Water content calibration coefficient B	0.404*
Water content calibration coefficient C	0.02852*
Ion content calibration coefficient A	1
Ion content calibration coefficient B	1
Ion content calibration coefficient C	0

* Manufacturer default values (Sentek 2003).

7.2 RS-232 Interface and Other Assembly Components

No configuration is required.



8 ASSEMBLY, SOIL WATER CONTENT PROFILE, POPULATION, 2 METER, 1 SENSOR COMMAND AND CONTROL (CF00810010)

There is no command and control for this assembly.

8.1 Error Handling

N/A

8.2 Sensor Controls Specification

N/A



<i>Title:</i> NEON Sensor Command, Control and Configuration (C3) Document: Soil Water Content and Soil Water Ion Content		<i>Date:</i> 05/26/2022
<i>NEON Doc. #:</i> NEON.DOC.000613	<i>Author:</i> E. Ayres	<i>Revision:</i> E

9 ASSEMBLY, SOIL WATER CONTENT PROFILE, POPULATION, 2 METER, 2 SENSORS INTRODUCTION (CF00810020)

Soil Water Content Profile subsystem data product name and number are Soil water content and salinity and NEON.DOM.SIT.DP0.00094, respectively. A description of how sensor readings shall be converted to the data product is presented in the associated ATBD (AD[03]).



10 ASSEMBLY, SOIL WATER CONTENT PROFILE, POPULATION, 2 METER, 2 SENSORS OVERVIEW OF SENSOR CONFIGURATION (CF00810020)

10.1 TriSCAN Sensors

Table 3. Sensor configuration settings.

Parameter	Default Setting
Acquisition rate: Soil water content	0.1 Hz
Acquisition rate: Soil water ion content	0.1 Hz
Water content high normalization value ($SWCF_{air}$)	#CVAL
Water content low normalization value ($SWCF_{water}$)	#CVAL
Water ion content high normalization value ($SWICF_{air}$)	#CVAL
Water ion content low normalization value ($SWICF_{water}$)	#CVAL
Water content calibration coefficient A	0.1957*
Water content calibration coefficient B	0.404*
Water content calibration coefficient C	0.02852*
Ion content calibration coefficient A	1
Ion content calibration coefficient B	1
Ion content calibration coefficient C	0

* Manufacturer default values (Sentek 2003).

10.2 RS-232 Interface and Other Assembly Components

No configuration is required.



11 ASSEMBLY, SOIL WATER CONTENT PROFILE, POPULATION, 2 METER, 2 SENSORS COMMAND AND CONTROL (CF00810020)

There is no command and control for this assembly.

11.1 Error Handling

N/A

11.2 Sensor Controls Specification

N/A



<i>Title:</i> NEON Sensor Command, Control and Configuration (C3) Document: Soil Water Content and Soil Water Ion Content		<i>Date:</i> 05/26/2022
<i>NEON Doc. #:</i> NEON.DOC.000613	<i>Author:</i> E. Ayres	<i>Revision:</i> E

12 ASSEMBLY, SOIL WATER CONTENT PROFILE, POPULATION, 2 METER, 3 SENSORS INTRODUCTION (CF00810030)

Soil Water Content Profile subsystem data product name and number are Soil water content and salinity and NEON.DOM.SIT.DP0.00094, respectively. A description of how sensor readings shall be converted to the data product is presented in the associated ATBD (AD[03]).



13 ASSEMBLY, SOIL WATER CONTENT PROFILE, POPULATION, 2 METER, 3 SENSORS OVERVIEW OF SENSOR CONFIGURATION (CF00810030)

13.1 TriSCAN Sensors

Table 4. Sensor configuration settings.

Parameter	Default Setting
Acquisition rate: Soil water content	0.1 Hz
Acquisition rate: Soil water ion content	0.1 Hz
Water content high normalization value ($SWCF_{air}$)	#CVAL
Water content low normalization value ($SWCF_{water}$)	#CVAL
Water ion content high normalization value ($SWICF_{air}$)	#CVAL
Water ion content low normalization value ($SWICF_{water}$)	#CVAL
Water content calibration coefficient A	0.1957*
Water content calibration coefficient B	0.404*
Water content calibration coefficient C	0.02852*
Ion content calibration coefficient A	1
Ion content calibration coefficient B	1
Ion content calibration coefficient C	0

* Manufacturer default values (Sentek 2003).

13.2 RS-232 Interface and Other Assembly Components

No configuration is required.



14 ASSEMBLY, SOIL WATER CONTENT PROFILE, POPULATION, 2 METER, 3 SENSORS COMMAND AND CONTROL (CF00810030)

There is no command and control for this assembly.

14.1 Error Handling

N/A

14.2 Sensor Controls Specification

N/A



<i>Title:</i> NEON Sensor Command, Control and Configuration (C3) Document: Soil Water Content and Soil Water Ion Content		<i>Date:</i> 05/26/2022
<i>NEON Doc. #:</i> NEON.DOC.000613	<i>Author:</i> E. Ayres	<i>Revision:</i> E

15 ASSEMBLY, SOIL WATER CONTENT PROFILE, POPULATION, 2 METER, 4 SENSORS INTRODUCTION (CF00810040)

Soil Water Content Profile subsystem data product name and number are Soil water content and salinity and NEON.DOM.SIT.DP0.00094, respectively. A description of how sensor readings shall be converted to the data product is presented in the associated ATBD (AD[03]).



16 ASSEMBLY, SOIL WATER CONTENT PROFILE, POPULATION, 2 METER, 4 SENSORS OVERVIEW OF SENSOR CONFIGURATION (CF00810040)

16.1 TriSCAN Sensors

Table 5. Sensor configuration settings.

Parameter	Default Setting
Acquisition rate: Soil water content	0.1 Hz
Acquisition rate: Soil water ion content	0.1 Hz
Water content high normalization value ($SWCF_{air}$)	#CVAL
Water content low normalization value ($SWCF_{water}$)	#CVAL
Water ion content high normalization value ($SWICF_{air}$)	#CVAL
Water ion content low normalization value ($SWICF_{water}$)	#CVAL
Water content calibration coefficient A	0.1957*
Water content calibration coefficient B	0.404*
Water content calibration coefficient C	0.02852*
Ion content calibration coefficient A	1
Ion content calibration coefficient B	1
Ion content calibration coefficient C	0

* Manufacturer default values (Sentek 2003).

16.2 RS-232 Interface and Other Assembly Components

No configuration is required.



17 ASSEMBLY, SOIL WATER CONTENT PROFILE, POPULATION, 2 METER, 4 SENSORS COMMAND AND CONTROL (CF00810040)

There is no command and control for this assembly.

17.1 Error Handling

N/A

17.2 Sensor Controls Specification

N/A



<i>Title:</i> NEON Sensor Command, Control and Configuration (C3) Document: Soil Water Content and Soil Water Ion Content		<i>Date:</i> 05/26/2022
<i>NEON Doc. #:</i> NEON.DOC.000613	<i>Author:</i> E. Ayres	<i>Revision:</i> E

18 ASSEMBLY, SOIL WATER CONTENT PROFILE, POPULATION, 2 METER, 6 SENSORS INTRODUCTION (CF00810060)

Soil Water Content Profile subsystem data product name and number are Soil water content and salinity and NEON.DOM.SIT.DP0.00094, respectively. A description of how sensor readings shall be converted to the data product is presented in the associated ATBD (AD[03]).



19 ASSEMBLY, SOIL WATER CONTENT PROFILE, POPULATION, 2 METER, 6 SENSORS OVERVIEW OF SENSOR CONFIGURATION (CF00810060)

19.1 TriSCAN Sensors

Table 6. Sensor configuration settings.

Parameter	Default Setting
Acquisition rate: Soil water content	0.1 Hz
Acquisition rate: Soil water ion content	0.1 Hz
Water content high normalization value ($SWCF_{air}$)	#CVAL
Water content low normalization value ($SWCF_{water}$)	#CVAL
Water ion content high normalization value ($SWICF_{air}$)	#CVAL
Water ion content low normalization value ($SWICF_{water}$)	#CVAL
Water content calibration coefficient A	0.1957*
Water content calibration coefficient B	0.404*
Water content calibration coefficient C	0.02852*
Ion content calibration coefficient A	1
Ion content calibration coefficient B	1
Ion content calibration coefficient C	0

* Manufacturer default values (Sentek 2003).

19.2 RS-232 Interface and Other Assembly Components

No configuration is required.



20 ASSEMBLY, SOIL WATER CONTENT PROFILE, POPULATION, 2 METER, 6 SENSORS COMMAND AND CONTROL (CF00810060)

There is no command and control for this assembly.

20.1 Error Handling

N/A

20.2 Sensor Controls Specification

N/A



<i>Title:</i> NEON Sensor Command, Control and Configuration (C3) Document: Soil Water Content and Soil Water Ion Content		<i>Date:</i> 05/26/2022
<i>NEON Doc. #:</i> NEON.DOC.000613	<i>Author:</i> E. Ayres	<i>Revision:</i> E

21 ASSEMBLY, SOIL WATER CONTENT PROFILE, POPULATION, 3 METER INTRODUCTION (CF00813000)

Soil Water Content Profile subsystem data product name and number are Soil water content and salinity and NEON.DOM.SIT.DP0.00094, respectively. A description of how sensor readings shall be converted to the data product is presented in the associated ATBD (AD[03]).



22 ASSEMBLY, SOIL WATER CONTENT PROFILE, POPULATION, 3 METER OVERVIEW OF SENSOR CONFIGURATION (CF00813000)

22.1 TriSCAN Sensors

Table 7. Sensor configuration settings.

Parameter	Default Setting
Acquisition rate: Soil water content	0.1 Hz
Acquisition rate: Soil water ion content	0.1 Hz
Water content high normalization value ($SWCF_{air}$)	#_CVAL
Water content low normalization value ($SWCF_{water}$)	#_CVAL
Water ion content high normalization value ($SWICF_{air}$)	#_CVAL
Water ion content low normalization value ($SWICF_{water}$)	#_CVAL
Water content calibration coefficient A	0.1957*
Water content calibration coefficient B	0.404*
Water content calibration coefficient C	0.02852*
Ion content calibration coefficient A	1
Ion content calibration coefficient B	1
Ion content calibration coefficient C	0

* Manufacturer default values (Sentek 2003).

22.2 RS-232 Interface and Other Assembly Components

No configuration is required.



23 ASSEMBLY, SOIL WATER CONTENT PROFILE, POPULATION, 3 METER COMMAND AND CONTROL (CF00813000)

There is no command and control for this assembly.

23.1 Error Handling

N/A

23.2 Sensor Controls Specification

N/A



<i>Title:</i> NEON Sensor Command, Control and Configuration (C3) Document: Soil Water Content and Soil Water Ion Content		<i>Date:</i> 05/26/2022
<i>NEON Doc. #:</i> NEON.DOC.000613	<i>Author:</i> E. Ayres	<i>Revision:</i> E

24 ASSEMBLY INTEGRATION

The shallowest TriSCAN sensors in the profile, or only sensor if the profile consists of one sensor, produces two data streams (Stream:00 : raw volumetric soil water content and Stream:01 : raw volumetric soil ion content). The 2nd shallowest sensor produces the same data streams, which are assigned to the next sequential stream number (i.e., Stream:02 : raw volumetric soil water content and Stream:03 : raw volumetric soil ion content). If additional deeper sensors are present they produce the same data streams, which are assigned sequentially (shallowest to deepest) to the next stream number. The maximum number of TriSCAN sensors within a profile is limited to 8, therefore, the highest stream number will be Stream:15.



25 APPENDIX

25.1 List of Level 0 Data Product

Table 8. List of Level 0 data product associated with DPName: Soil water content and salinity



neon
Operated by Battelle

Title: NEON Sensor Command, Control and Configuration (C3) Document: Soil Water Content and Soil Water Ion Content

Date: 05/26/2022

NEON Doc. #: NEON.DOC.000613

Author: E. Ayres

Revision: E

DGD Agile PN	DPNumber	fieldName	description	Acquisition frequency (Hz)	dataType	units
CF00810000	NEON.DOM.SITE.DP0.00094.001.02081.HOR.VER.000	rawVSWC0	Raw measurement of volumetric soil water content from shallowest depth	0.1	real	percent
	NEON.DOM.SITE.DP0.00094.001.02082.HOR.VER.000	rawVSIC0	Raw measurement of volumetric soil ion content from shallowest depth	0.1	real	NA
	NEON.DOM.SITE.DP0.00094.001.02083.HOR.VER.000	rawVSWC1	Raw measurement of volumetric soil water content from 2nd shallowest depth	0.1	real	percent
	NEON.DOM.SITE.DP0.00094.001.02084.HOR.VER.000	rawVSIC1	Raw measurement of volumetric soil ion content from 2nd shallowest depth	0.1	real	NA
	NEON.DOM.SITE.DP0.00094.001.02085.HOR.VER.000	rawVSWC2	Raw measurement of volumetric soil water content from 3rd shallowest depth	0.1	real	percent
	NEON.DOM.SITE.DP0.00094.001.02086.HOR.VER.000	rawVSIC2	Raw measurement of volumetric soil ion content from 3rd shallowest depth	0.1	real	NA
	NEON.DOM.SITE.DP0.00094.001.02087.HOR.VER.000	rawVSWC3	Raw measurement of volumetric soil water content from 4th shallowest depth	0.1	real	percent
	NEON.DOM.SITE.DP0.00094.001.02088.HOR.VER.000	rawVSIC3	Raw measurement of volumetric soil ion content from 4th shallowest depth	0.1	real	NA
	NEON.DOM.SITE.DP0.00094.001.02089.HOR.VER.000	rawVSWC4	Raw measurement of volumetric soil water content from 5th shallowest depth	0.1	real	percent
	NEON.DOM.SITE.DP0.00094.001.02090.HOR.VER.000	rawVSIC4	Raw measurement of volumetric soil ion content from 5th shallowest depth	0.1	real	NA



neon
Operated by Battelle

Title: NEON Sensor Command, Control and Configuration (C3) Document: Soil Water Content and Soil Water Ion Content

Date: 05/26/2022

NEON Doc. #: NEON.DOC.000613

Author: E. Ayres

Revision: E

	NEON.DOM.SITE.DP0.00094.001.02091.HOR.VER.000	rawVSWC5	Raw measurement of volumetric soil water content from 6th shallowest depth	0.1	real	percent
	NEON.DOM.SITE.DP0.00094.001.02092.HOR.VER.000	rawVSIC5	Raw measurement of volumetric soil ion content from 6th shallowest depth	0.1	real	NA
	NEON.DOM.SITE.DP0.00094.001.02093.HOR.VER.000	rawVSWC6	Raw measurement of volumetric soil water content from 7th shallowest depth	0.1	real	percent
	NEON.DOM.SITE.DP0.00094.001.02094.HOR.VER.000	rawVSIC6	Raw measurement of volumetric soil ion content from 7th shallowest depth	0.1	real	NA
	NEON.DOM.SITE.DP0.00094.001.02095.HOR.VER.000	rawVSWC7	Raw measurement of volumetric soil water content from 8th shallowest depth	0.1	real	percent
	NEON.DOM.SITE.DP0.00094.001.02096.HOR.VER.000	rawVSIC7	Raw measurement of volumetric soil ion content from 8th shallowest depth	0.1	real	NA
CF00810010	NEON.DOM.SITE.DP0.00094.001.02081.HOR.VER.000	rawVSWC0	Raw measurement of volumetric soil water content from shallowest depth	0.1	real	percent
	NEON.DOM.SITE.DP0.00094.001.02082.HOR.VER.000	rawVSIC0	Raw measurement of volumetric soil ion content from shallowest depth	0.1	real	NA
CF00810020	NEON.DOM.SITE.DP0.00094.001.02081.HOR.VER.000	rawVSWC0	Raw measurement of volumetric soil water content from shallowest depth	0.1	real	percent
	NEON.DOM.SITE.DP0.00094.001.02082.HOR.VER.000	rawVSIC0	Raw measurement of volumetric soil ion content from shallowest depth	0.1	real	NA
	NEON.DOM.SITE.DP0.00094.001.02083.HOR.VER.000	rawVSWC1	Raw measurement of volumetric soil water content from 2nd shallowest depth	0.1	real	percent



neon
Operated by Battelle

Title: NEON Sensor Command, Control and Configuration (C3) Document: Soil Water Content and Soil Water Ion Content

Date: 05/26/2022

NEON Doc. #: NEON.DOC.000613

Author: E. Ayres

Revision: E

	NEON.DOM.SITE.DP0.00094.001.02084.HOR.VER.000	rawVSIC1	Raw measurement of volumetric soil ion content from 2nd shallowest depth	0.1	real	NA
CF00810030	NEON.DOM.SITE.DP0.00094.001.02081.HOR.VER.000	rawVSWC0	Raw measurement of volumetric soil water content from shallowest depth	0.1	real	percent
	NEON.DOM.SITE.DP0.00094.001.02082.HOR.VER.000	rawVSICO	Raw measurement of volumetric soil ion content from shallowest depth	0.1	real	NA
	NEON.DOM.SITE.DP0.00094.001.02083.HOR.VER.000	rawVSWC1	Raw measurement of volumetric soil water content from 2nd shallowest depth	0.1	real	percent
	NEON.DOM.SITE.DP0.00094.001.02084.HOR.VER.000	rawVSIC1	Raw measurement of volumetric soil ion content from 2nd shallowest depth	0.1	real	NA
	NEON.DOM.SITE.DP0.00094.001.02085.HOR.VER.000	rawVSWC2	Raw measurement of volumetric soil water content from 3rd shallowest depth	0.1	real	percent
	NEON.DOM.SITE.DP0.00094.001.02086.HOR.VER.000	rawVSIC2	Raw measurement of volumetric soil ion content from 3rd shallowest depth	0.1	real	NA
CF00810040	NEON.DOM.SITE.DP0.00094.001.02081.HOR.VER.000	rawVSWC0	Raw measurement of volumetric soil water content from shallowest depth	0.1	real	percent
	NEON.DOM.SITE.DP0.00094.001.02082.HOR.VER.000	rawVSICO	Raw measurement of volumetric soil ion content from shallowest depth	0.1	real	NA
	NEON.DOM.SITE.DP0.00094.001.02083.HOR.VER.000	rawVSWC1	Raw measurement of volumetric soil water content from 2nd shallowest depth	0.1	real	percent
	NEON.DOM.SITE.DP0.00094.001.02084.HOR.VER.000	rawVSIC1	Raw measurement of volumetric soil ion content from 2nd shallowest depth	0.1	real	NA



neon
Operated by Battelle

Title: NEON Sensor Command, Control and Configuration (C3) Document: Soil Water Content and Soil Water Ion Content

Date: 05/26/2022

NEON Doc. #: NEON.DOC.000613

Author: E. Ayres

Revision: E

	NEON.DOM.SITE.DP0.00094.001.02085.HOR.VER.000	rawVSWC2	Raw measurement of volumetric soil water content from 3rd shallowest depth	0.1	real	percent
	NEON.DOM.SITE.DP0.00094.001.02086.HOR.VER.000	rawVSIC2	Raw measurement of volumetric soil ion content from 3rd shallowest depth	0.1	real	NA
	NEON.DOM.SITE.DP0.00094.001.02087.HOR.VER.000	rawVSWC3	Raw measurement of volumetric soil water content from 4th shallowest depth	0.1	real	percent
	NEON.DOM.SITE.DP0.00094.001.02088.HOR.VER.000	rawVSIC3	Raw measurement of volumetric soil ion content from 4th shallowest depth	0.1	real	NA
CF00810060	NEON.DOM.SITE.DP0.00094.001.02081.HOR.VER.000	rawVSWC0	Raw measurement of volumetric soil water content from shallowest depth	0.1	real	percent
	NEON.DOM.SITE.DP0.00094.001.02082.HOR.VER.000	rawVSIC0	Raw measurement of volumetric soil ion content from shallowest depth	0.1	real	NA
	NEON.DOM.SITE.DP0.00094.001.02083.HOR.VER.000	rawVSWC1	Raw measurement of volumetric soil water content from 2nd shallowest depth	0.1	real	percent
	NEON.DOM.SITE.DP0.00094.001.02084.HOR.VER.000	rawVSIC1	Raw measurement of volumetric soil ion content from 2nd shallowest depth	0.1	real	NA
	NEON.DOM.SITE.DP0.00094.001.02085.HOR.VER.000	rawVSWC2	Raw measurement of volumetric soil water content from 3rd shallowest depth	0.1	real	percent
	NEON.DOM.SITE.DP0.00094.001.02086.HOR.VER.000	rawVSIC2	Raw measurement of volumetric soil ion content from 3rd shallowest depth	0.1	real	NA
	NEON.DOM.SITE.DP0.00094.001.02087.HOR.VER.000	rawVSWC3	Raw measurement of volumetric soil water content from 4th shallowest depth	0.1	real	percent



neon
Operated by Battelle

Title: NEON Sensor Command, Control and Configuration (C3) Document: Soil Water Content and Soil Water Ion Content

Date: 05/26/2022

NEON Doc. #: NEON.DOC.000613

Author: E. Ayres

Revision: E

	NEON.DOM.SITE.DP0.00094.001.02088.HOR.VER.000	rawVSIC3	Raw measurement of volumetric soil ion content from 4th shallowest depth	0.1	real	NA
	NEON.DOM.SITE.DP0.00094.001.02089.HOR.VER.000	rawVSWC4	Raw measurement of volumetric soil water content from 5th shallowest depth	0.1	real	percent
	NEON.DOM.SITE.DP0.00094.001.02090.HOR.VER.000	rawVSIC4	Raw measurement of volumetric soil ion content from 5th shallowest depth	0.1	real	NA
	NEON.DOM.SITE.DP0.00094.001.02091.HOR.VER.000	rawVSWC5	Raw measurement of volumetric soil water content from 6th shallowest depth	0.1	real	percent
	NEON.DOM.SITE.DP0.00094.001.02092.HOR.VER.000	rawVSIC5	Raw measurement of volumetric soil ion content from 6th shallowest depth	0.1	real	NA
	NEON.DOM.SITE.DP0.00094.001.02081.HOR.VER.000	rawVSWC0	Raw measurement of volumetric soil water content from shallowest depth	0.1	real	percent
CF00813000	NEON.DOM.SITE.DP0.00094.001.02082.HOR.VER.000	rawVSIC0	Raw measurement of volumetric soil ion content from shallowest depth	0.1	real	NA
	NEON.DOM.SITE.DP0.00094.001.02083.HOR.VER.000	rawVSWC1	Raw measurement of volumetric soil water content from 2nd shallowest depth	0.1	real	percent
	NEON.DOM.SITE.DP0.00094.001.02084.HOR.VER.000	rawVSIC1	Raw measurement of volumetric soil ion content from 2nd shallowest depth	0.1	real	NA
	NEON.DOM.SITE.DP0.00094.001.02085.HOR.VER.000	rawVSWC2	Raw measurement of volumetric soil water content from 3rd shallowest depth	0.1	real	percent
	NEON.DOM.SITE.DP0.00094.001.02086.HOR.VER.000	rawVSIC2	Raw measurement of volumetric soil ion content from 3rd shallowest depth	0.1	real	NA



neon
Operated by Battelle

Title: NEON Sensor Command, Control and Configuration (C3) Document: Soil Water Content and Soil Water Ion Content

Date: 05/26/2022

NEON Doc. #: NEON.DOC.000613

Author: E. Ayres

Revision: E

	NEON.DOM.SITE.DP0.00094.001.02087.HOR.VER.000	rawVSWC3	Raw measurement of volumetric soil water content from 4th shallowest depth	0.1	real	percent
	NEON.DOM.SITE.DP0.00094.001.02088.HOR.VER.000	rawVSIC3	Raw measurement of volumetric soil ion content from 4th shallowest depth	0.1	real	NA
	NEON.DOM.SITE.DP0.00094.001.02089.HOR.VER.000	rawVSWC4	Raw measurement of volumetric soil water content from 5th shallowest depth	0.1	real	percent
	NEON.DOM.SITE.DP0.00094.001.02090.HOR.VER.000	rawVSIC4	Raw measurement of volumetric soil ion content from 5th shallowest depth	0.1	real	NA
	NEON.DOM.SITE.DP0.00094.001.02091.HOR.VER.000	rawVSWC5	Raw measurement of volumetric soil water content from 6th shallowest depth	0.1	real	percent
	NEON.DOM.SITE.DP0.00094.001.02092.HOR.VER.000	rawVSIC5	Raw measurement of volumetric soil ion content from 6th shallowest depth	0.1	real	NA
	NEON.DOM.SITE.DP0.00094.001.02093.HOR.VER.000	rawVSWC6	Raw measurement of volumetric soil water content from 7th shallowest depth	0.1	real	percent
	NEON.DOM.SITE.DP0.00094.001.02094.HOR.VER.000	rawVSIC6	Raw measurement of volumetric soil ion content from 7th shallowest depth	0.1	real	NA
	NEON.DOM.SITE.DP0.00094.001.02095.HOR.VER.000	rawVSWC7	Raw measurement of volumetric soil water content from 8th shallowest depth	0.1	real	percent
	NEON.DOM.SITE.DP0.00094.001.02096.HOR.VER.000	rawVSIC7	Raw measurement of volumetric soil ion content from 8th shallowest depth	0.1	real	NA



neon
Operated by Battelle

Title: NEON Sensor Command, Control and Configuration (C3) Document: Soil Water Content and Soil Water Ion Content

Date: 05/26/2022

NEON Doc. #: NEON.DOC.000613

Author: E. Ayres

Revision: E

25.2 Assembly Schematic Drawing

N/A



26 BIBLIOGRAPHY

Sentek. 2003. TriSCAN Agronomic User Manual, Version 1.2a. Sentek, Stepney, South Australia.