

<i>Title:</i> NEON Sensor Configuration and Command and Control – 2D Wind Sonic Anemometer	<i>Author:</i> J. Roberti	<i>Date:</i> 05/28/2013
<i>NEON Doc. #:</i> NEON.DOC.000387		<i>Revision:</i> B


## NEON Sensor Configuration, Command and Control – 2D Wind Sonic Anemometer

<b>PREPARED BY:</b>	<b>ORGANIZATION:</b>	<b>DATE:</b>
Josh Roberti	FIU	05/28/2013

<b>APPROVALS (Name):</b>	<b>ORGANIZATION:</b>	<b>APPROVAL DATE:</b>
Hanne Buur	CCB DIR SE	05/25/2013
Hank Loescher	FIU	03/13/2013

<b>RELEASED BY (Name):</b>	<b>ORGANIZATION:</b>	<b>RELEASE DATE:</b>
Stephen Craft	CCB Admin	05/28/2013

See Configuration Management System for Approval History.



<i>Title:</i> NEON Sensor Configuration and Command and Control – 2D Wind Sonic Anemometer	<i>Author:</i> J. Roberti	<i>Date:</i> 05/28/2013
<i>NEON Doc. #:</i> NEON.DOC.000387		<i>Revision:</i> B

### Change Record

<b>REVISION</b>	<b>DATE</b>	<b>ECO #</b>	<b>DESCRIPTION OF CHANGE</b>
A	05/17/2012	ECO-00418	INITIAL RELEASE
B	05/28/2013	ECO-00593	Minor Updates

Title: NEON Sensor Configuration and Command and Control – 2D Wind Sonic Anemometer	Author: J. Roberti	Date: 05/28/2013
NEON Doc. #: NEON.DOC.000387		Revision: B

**TABLE OF CONTENTS**

1 DESCRIPTION..... 1

    1.1 Purpose ..... 1

    1.2 Scope..... 1

2 Related documents and acronyms ..... 2

    2.1 Applicable Documents ..... 2

    2.2 Reference Documents..... 2

    2.3 Acronyms ..... 2

    2.4 Verb Convention ..... 2

3 Introduction ..... 3

4 Overview of Sensor configuration ..... 3

5 Command and Control..... 5

    5.1 Error handling ..... 5

6 Appendix ..... 6

    6.1 Sensor error codes as of October 2011 ..... 6

**LIST OF TABLES**

**Table 1.** Sensor configuration settings. .... 4

**Table A1.** Error codes and interpretation of sensor error messages as of October 2011..... 6

<i>Title:</i> NEON Sensor Configuration and Command and Control – 2D Wind Sonic Anemometer	<i>Author:</i> J. Roberti	<i>Date:</i> 05/28/2013
<i>NEON Doc. #:</i> NEON.DOC.000387		<i>Revision:</i> B

## 1 DESCRIPTION

### 1.1 Purpose

This document specifies the configuration and command and control details for the 2D wind sonic anemometer sensors. It includes a detailed discussion of all necessary requirements for operational control parameters, conditions/constraints, set points, and any necessary error handling.

### 1.2 Scope

Gill’s WindObserver II 2D Sonic Anemometer (RD [03]), and Extreme Weather WindObserver 2D Sonic Anemometer Heated (RD [04]) will be used throughout NEON’s Observatory to monitor wind conditions. For the remainder of this document, the WindObserver II and Extreme Weather WindObserver will be referenced as ‘WOII’ and ‘EWWO’, respectively. Two versions of the WOII will be used: version 1 (no heater, NEON P/N: 0303440000) and version 3 (heater, NEON P/N: 0303440001). Version 1 will be used in locations where icing is not plausible, while version 3 will be used in areas where icing is likely. If severe icing is eminent (e.g. Tundra, mountainous terrain), the EWWO (NEON P/N: 0303440005) will be used. The firmware that shall be used with these sensors is version 6.01. This document specifies the configuration and command and control that are needed for operating each of these anemometers. It does not provide implementation details, except for cases where these stem directly from the sensor conditions as described here.

Title: NEON Sensor Configuration and Command and Control – 2D Wind Sonic Anemometer	Author: J. Roberti	Date: 05/28/2013
NEON Doc. #: NEON.DOC.000387		Revision: B

## 2 RELATED DOCUMENTS AND ACRONYMS

### 2.1 Applicable Documents

AD [01]	NEON.DOC.000001	NEON Observatory Design (NOD) Requirements
AD [02]	NEON.DOC.000291	NEON Configured Sensor List
AD [03]	NEON.DOC.005003	NEON Scientific Data Products Catalog
AD [04]	NEON.DOC.005005	NEON Level 0 Data Products Catalog

### 2.2 Reference Documents

RD [01]	NEON.DOC.000008	NEON Acronym List
RD [02]	NEON.DOC.000243	NEON Glossary of Terms
RD [03]	Gill Instruments Ltd. 2007. User Manual: WindObserver II Ultrasonic Anemometer. Document # 1390-PS-004. Issue 17.	
RD [04]	Gill Instruments Ltd. 2011. User Manual: Extreme Weather WindObserver with Enhanced Heating Ultrasonic Anemometer. Document # 1390-PS-0018. Issue 6.	
RD [05]	Sims, Murree. "RE: Questions regarding Gill's 2D Ultrasonic anemometers." Email to Joshua Roberti. 14 March 2012. (P:\FIU\TIS Assemblies\4. Tower Mid Level Wind-2D sonic\Questions regarding Gill's 2D Ultrasonic anemometers.pdf)	

### 2.3 Acronyms

Acronym	Explanation
ATBD	Algorithm Theoretical Basis Document
C <sup>3</sup>	Command, Control, and Configuration Document
SOP	Standard Operating Procedures
QA/QC	Quality Assurance/Quality Control
TIS	Terrestrial Instrument System
L0	Level 0
L1	Level 1
ENG	NEON Engineering group
CI	NEON Cyberinfrastructure group
DPS	NEON Data Products group
CVAL	NEON Calibration, Validation, and Audit Laboratory
WOII	Gill's WindObserver II
EWWO	Gill's Extreme Weather WindObserver
P/N	Part Number

### 2.4 Verb Convention

"Shall" is used whenever a statement expresses a convention that is binding. The verbs "should" and "may" express non-mandatory provisions. "Will" is used to express a declaration of purpose on the part of the design activity.

<i>Title:</i> NEON Sensor Configuration and Command and Control – 2D Wind Sonic Anemometer	<i>Author:</i> J. Roberti	<i>Date:</i> 05/28/2013
<i>NEON Doc. #:</i> NEON.DOC.000387		<i>Revision:</i> B

### 3 INTRODUCTION

This document describes the configuration and command and control related with each of the 2D wind sonic anemometer sensors and their corresponding data products (FIU.0.0001). The communication standard for digital output with the 2D wind sonic anemometer shall be RS422. For more information regarding maintenance or topics concerning computer algorithms, please refer to the SOP (AD [#]) and ATBD (AD [#]) documents, respectively.

### 4 OVERVIEW OF SENSOR CONFIGURATION

Three ultrasonic anemometer models will be used throughout NEON. Although there is some discrepancy, the default configurations of both models are mostly consistent with respect to one another. In the attempt to obtain the highest quality of data and ensure uniformity across the Observatory, some of the default settings of each anemometer model shall be changed (Table 1).

Title: NEON Sensor Configuration and Command and Control – 2D Wind Sonic Anemometer	Author: J. Roberti	Date: 05/28/2013
NEON Doc. #: NEON.DOC.000387		Revision: B

**Table 1.** Sensor configuration settings. Certain configurations are only applicable for the WOII model; these configurations are denoted by \*. If a sensor is equipped with a heater (WOII version 3 or EWWO), the heater shall be activated (H2) prior to deployment within the field. This adapted setting is denoted by \*\*. If the sensor is not equipped with a heater, the default setting shall not be changed.

Parameter	Default Setting	Code	Adapted Setting	Code
Baud rate	9600 (pulses s <sup>-1</sup> )	B3		
Duplex mode	Full Duplex	E1		
Data & parity options	8 bit, no parity, 1 stop bit	F1		
Averaging	Off	G0000		
Heater	Off	H1	Activated**	H2**
NMEA settings	"IIMWV"	K1		
ASCII message terminator	"CR LF"	L1		
Data output	ASCII Polar continuous	M2	ASCII UV continuous	M1
Node address	<A>	NA		
Output format	CSV	O1		
Output (acquisition) rate	1 Hz	P1		
Speed output	m s <sup>-1</sup>	U1		
Vertical output padding	Off	V1		
45° Offset (WOII)	Off	X1	Mount Upside Down	X3
45° Offset (EWWO)	On	X2	Mount Upside Down	X3
Speed of sound / Sonic Temp.*	Off	A0	On	A3
Wind wraparound*	On	C1	Off	C2
Data acquisition streams	Start of string Unit Identification U-component V-component Units Speed of sound * Sonic temperature * Status End of string Check sum	N/A		

Title: NEON Sensor Configuration and Command and Control – 2D Wind Sonic Anemometer	Author: J. Roberti	Date: 05/28/2013
NEON Doc. #: NEON.DOC.000387		Revision: B

## 5 COMMAND AND CONTROL

### 5.1 Error handling

If the sensor is operating under the influence of error, the data output’s status will reference a variable other than 00, 60, or 66, and the condition shall be deemed “failed” (Table A1). If a failed condition lasts for more than three hours, it is recommended that the issue be forwarded to Problem Resolution and Tracking. Since Gill Instruments does not offer any command / control functions to handle errors, the sensor may need to be manually repaired.

For all purposes, if the anemometer is equipped with a heater, the setting of the heater shall be changed from its default of ‘off’ (H1) to the adapted setting of ‘activated’. The heater of the WOII will automatically be turned on when the *temperature of the transducer* is  $<15^{\circ}\text{C}$ , and will turn off when the *temperature of the transducer* is  $\geq 25^{\circ}\text{C}$ . The heater of the EWWO will automatically be turned on when the *ambient air temperature* is  $\leq 5^{\circ}\text{C}$  and will turn off when the *ambient air temperature* is  $\geq 10^{\circ}\text{C}$  (RD [05]).

Since the heaters for each model operate automatically, there is no need for any heating command / control statement at this time. Although it may be beneficial to generate a command /control function that is based on site-specific climate characteristics, it is not possible to override the preset heater settings (RD [05]).



Title: NEON Sensor Configuration and Command and Control – 2D Wind Sonic Anemometer	Author: J. Roberti	Date: 05/28/2013
NEON Doc. #: NEON.DOC.000387		Revision: B

**6 APPENDIX**

**6.1 Sensor error codes as of October 2011**

**Table A1.** Error codes and interpretation of sensor error messages as of October 2011 (RD [03], RD [04]).

Code	Status	Condition
00	OK Heating set off	Valid data, sufficient samples in average period
60 (H2) 66 (H3)	OK and heating enabled	Valid data, sufficient samples in average period
A	OK	NMEA data Acceptable
01	Axis 1 failed	Insufficient samples in average period on U axis
02	Axis 2 failed	Insufficient samples in average period on V axis
04	Axis 1 and 2 failed	Insufficient samples in average period on both axes
08	NVM error	NVM checksum failed, data could be uncalibrated.
09	ROM error	ROM checksum failed, data could be uncalibrated.
10	System gain at max.	Inaccurate data likely
50	Marginal system gain	Data valid, but marginal operation
51	Measurement average building.	Data valid but warns that average period not reached when averaging used.
62 (H2) 67 (H3)	No power to heating module	Valid data still output
63 (H2) 68 (H3)	Hardware fault – e.g Heater pcb faulty	Valid data still output.
65 (H2) 69 (H3)	Warning – Heater Supply volts too high or pcb too hot.	Valid data still output.
V	NMEA data Void	Invalid data output